



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 1, 2011 INITIALS: JEP OWNER: WASHINGTON STATE PATROL
DISTRICT: 3 ADDRESS:
LOCATION: FM SHOP COUNTY:
MAKE: HAENNI TYPE PORT. WHEEL SCALE ACCURACY CLASS: IIII
MODEL: 101 ID. NO: 27767 CAPACITY: 20000 lbs. SECTIONS:
MINIMUM GRADUATIONS: 50 lbs. REG. ELEMENT: DIAL PLATFORM 15"X26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	5000	5000	0				
CENTER	10000	10000	0				
CENTER	15000	15000	0				
CENTER	20000	19900	-100				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report and Certificate of Calibration",
Certification number(s): 36293-W-01

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Mar 28, 2011

Print name: DAVID CROMER

State of Washington
County of Thurston
Signed and sworn to (or affirmed before
me on Mar 28, 2011 by

Brenda A Biscay
(Signature of notary public)
Brenda A Biscay
Printed Name of Notary
My appointment expires 3.12.12

